

# Display Report

## Analysis Info

Analysis Name D:\Data\engle\CS017037\_GC3\_01\_8511.d  
Method 0303-MS-neg.m  
Sample Name CS017037  
Comment

Acquisition Date 8/5/2021 11:19:32 AM

Operator Demo User  
Instrument impact II 1825265.10256

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Negative	Set Nebulizer	0.4 Bar
Focus	Not active	Set Capillary	3000 V	Set Dry Heater	180 °C
Scan Begin	50 m/z	Set End Plate Offset	-500 V	Set Dry Gas	4.0 l/min
Scan End	1300 m/z	Set Charging Voltage	2000 V	Set Divert Valve	Source
		Set Corona	0 nA	Set APCI Heater	0 °C

